



Date Issued: 1/17/2024

RE: Vitalacy SmartBadge Radiation Exposure

This letter is to validate the radiation exposure level of the Vitalacy SmartBadge wearable device.

The C10 SmartBadge (Vitalacy SmartBadge) is FCC certified.

The C10 SmartBadge passed with a SAR of 0.3 W/kg. This is 5x less than the FCC limit for public exposure which is a SAR level of 1.6 W/kg. The SmartBadge (C10) is FCC certified to produce acceptable radiation exposure levels as defined by the US government to be safe for wearables.

As a comparison, the radiation exposure of the iPhone 12 is .98 W/kg.

Attached is the official SAR report for the C10 SmartBadge for reference.

Please contact Vitalacy using help@vitalacy.com for any further questions.